



**ICAR - INDIAN INSTITUTE OF RICE RESEARCH**  
(Formerly Directorate of Rice Research)  
Rajendranagar, Hyderabad – 30: Telangana  
Phone No's: 040-24591201: FAX: 040-24591217



F.No.6-I-44/Non Refrigerated Centrifuge/Evaluation of BAS 750 02 F...blast of Rice/Prakasam/P&S/16-17

Date:07-02-2017

o/c  
/3513

**BY REGD.POST**

**ENQUIRY NO: 113 (01-27)**

**TENDER FOR : "Non Refrigerated Centrifuge"**  
**DUE DATE : 27/02/2017**

To

**Sub: Procurement of Non Refrigerated Centrifuge – Reg.**

Dear Sir,

Tenders are invited for Procurement of **Non Refrigerated Centrifuge** as per the following specifications Copy Enclosed.

You are therefore, requested to kindly submit your Tender for the above said items to the Senior Administrative Officer, Indian Institute of Rice Research, Rajendranagar, Hyderabad 500 030 by Post/Courier.

**NOTE: LIMITED TENDERS BY POST ARE ONLY ACCEPTED**

P.T.O

**TERMS & CONDITIONS**

1. The last date for submission of Tenders is **27-02-2017 at 02:30 PM.**
1. Tenders should be submitted only by **POST / COURIER** in a sealed cover addressed to Senior Administration Officer, Indian Institute of Rice Research, Rajendranagar, Hyderabad – 500030 so as to reach by at 02:30 PM in the closing day. Delay on account of postal / other will not be the responsibly of the IIRR. The cover containing Tender should invariably be superscribed separately "**Tenders are invited for Non Refrigerated Centrifuge (1 No) .**
2. 2% of EMD drawn in favour of Director, IIRR, Rajendranagar, Hyderabad on estimated cost to be enclosed along with the Tender.
3. The successful bidder has to furnish an amount, equivalent to 10 % of the order value as security Deposit which will be refunded / returned only after satisfactory completion of the contractual obligations including warranty.

4. No advance payment will be made. Payment will be made through on-line after satisfactory installation/certification of the concerned Officer. Please furnish the following details:

Account Name :  
Account Number :  
Account Type :  
MICR Code :  
Branch :  
IFSC Code :  
Bank Address :

5. The rates quoted should be net payable for each item for delivery at the Institute at the address given above (inclusive of all taxes, packing forwarding, transport, insurance and excluding rebate/discount etc.)
6. This institute is not in a position to supply and "D" or "C" forms
7. While quoting the rates please mention the following:  
Approximate time for supply of stores from the date of placing order
8. Tender which does not conform to the above terms and conditions will not be considered. The Director, IIRR reserves the right to accept or reject any or all Tenders without assigning any reasons thereof.

Yours faithfully,




**Asst. Administrative Officer**

### 5. Non Refrigerated Centrifuge – Qty-1

The equipment should have following facilities viz.

- 24-Place refrigerated Micro centrifuge with two-knob operation.
- Very compact footprint to fit on every lab bench with very quiet operation and low access height for simple loading and unloading and unloading of samples
- SOFT brake functions for slow acceleration and braking.
- The system should be optimized for cell separation by gradient centrifugation
- At set RPM function, the timer should be started only when the selected rotational speed is reached for reproducible centrifugation runs
- Key lock facility to prevent unintentional adjustments
- With Opti bowl Design for Using machine without lid
- Max RPM 15000 and Operating power range: 220-240V
- Max RCF 21,130 X g
- Rotor capacity 24X 1.5/2.0 ml
- Separate Short Spin button for fast and short spins
- IVD conform
- Polytetrafluoroethylene (PTFE) Coating for increased chemical Resistance.
- Aerosol tight quick lock lid and polypropylene lid not preferable
- Should be provided with appropriate voltage stabilizer
- The equipment should have a minimum of 1 year warranty (Desirable: Warranty for 2 years)
- The quote should be accompanied by a point-by- point compliance statement for each item of specification and should be compulsorily supported by appropriate technical literature.

S. No	Specification item	Whether complied or not (Yes or No)	If Yes/No additional details and improved specifications, if any	Any additional remarks, if any

  
V. Prakasam  
PI of the Project