



**ICAR - INDIAN INSTITUTE OF RICE RESEARCH**  
(Formerly Directorate of Rice Research)  
Rajendranagar, Hyderabad – 30: Telangana  
Phone No's: 040-24591201: FAX: 040-24591217



F.No.6-1 -37/Precision Electronic Balance/DBT-Marker-assisted...in rice/RMS/P&S/16-17

3328

Date:20-01-2017

**BY REGD.POST**

**ENQUIRY NO: 111(01-31)**

**TENDER FOR : "Precision Electronic Balance"**  
**DUE DATE : 06/02/2017**

To

**Sub: Procurement of Precision Electronic Balance (1 No)– Reg.**

Dear Sir,

Tenders are invited for Procurement of Precision Electronic Balance (1 No) as per the following specifications Copy Enclosed.

You are therefore, requested to kindly submit your Tender for the above said items to the Senior Administrative Officer, Indian Institute of Rice Research, Rajendranagar, Hyderabad 500 030 by Post/Courier.

**NOTE: LIMITED TENDERS BY POST ARE ONLY ACCEPTED**

P.T.O

**TERMS & CONDITIONS**

1. The last date for submission of Tenders is **06-02-2017 at 02:30 PM.**
2. Tenders should be submitted only by **POST / COURIER** in a sealed cover addressed to Senior Administration Officer, Indian Institute of Rice Research, Rajendranagar, Hyderabad – 500030 so as to reach by at 02:30 PM in the closing day. Delay on account of postal / other will not be the responsibly of the IIRR. The cover containing Tender should invariably be superscribed separately **"Tenders are invited for Precision Electronic Balance (1 No)"**.
3. 2% of EMD drawn in favour of Director, IIRR, Rajendranagar, Hyderabad on estimated cost to be enclosed along with the Tender.
4. The successful bidder has to furnish an amount, equivalent to 10 % of the order value as security Deposit which will be refunded / returned only after satisfactory completion of the contractual obligations including warranty.

5. No advance payment will be made. Payment will be made through on-line after satisfactory installation/certification of the concerned Officer. Please furnish the following details:

Account Name :  
Account Number :  
Account Type :  
MICR Code :  
Branch :  
IFSC Code :  
Bank Address :

6. The rates quoted should be net payable for each item for delivery at the Institute at the address given above (inclusive of all taxes, packing forwarding, transport, insurance and excluding rebate/discount etc.)

7. This institute is not in a position to supply and "D" or "C" forms

8. While quoting the rates please mention the following:

Approximate time for supply of stores from the date of placing order

9. Tender which does not conform to the above terms and conditions will not be considered. The Director, IIRR reserves the right to accept or reject any or all Tenders without assigning any reasons thereof.

Yours faithfully,

*Rajeev*  
21/1/12

**Asst. Administrative Officer**

*JKC*  
20/1/12

**Copy to**

- Dr. B. Sailaja, Scientist, GIS – to be placed in IIRR Website/Central Procurement Portal

**Precision Electronic Balance (1 No.)**

S. No.	Specification Item	Features required
1	Automation Interface	Keys
2	Certifications/Compliance	ISO/GLP
3	Connection Type	RS232
4	Display-Type	Backlit LCD
5	Interface	Intuitive interface with clear readout: large numbers and backlit display
6	Readability and repeatability from	0.1 to 0.001g
7	Overload protection	Upto 10 Kgs
8	Low temperature drift effect	3.0ppm/degreeC
9	Setting time	Fast, <1.5 sec. settling time
10	Edges	Rounded edges and smooth surfaces for easy cleaning
11	Levelling of Front feet	Yes
12	Should comply with ISO/GLP documentation requirements	Yes
13	Electrical Requirements	240V 50/60Hz
14	Material	ABS Plastic, Die-cast Aluminum, Stainless Steel
15	Temperature Range (Metric)	10 to 30°C
16	Wattage	10w
17	Amperage	0.3A

18. A certified compliance statement (sample given below) wherein all the claims and specifications about the instrument should be supported by a printed and authentic technical literature.

S. No	Specification item	Whether complied or not (Yes or No)	If Yes/No additional details and improved specifications, if any	Any additional remarks, if any

*Handwritten signature*